

Tax Invoice

e-Invoice



IRN : b0964b1848159e0ba1ebbf8a7f2b31b8929fcbbbe1d76-
5f01df00eaa8efa74ba
Ack No. : 112422285111059
Ack Date : 18-Oct-24

MIRRANT AUTOMATION PRIVATE LIMITED PLOT NO:6A/5, IDA PHASE-I, PATANCHERU, HYDERABAD, TS (36) 502319 GSTIN/UID: 36AAKCM0817G1ZK State Name : Telangana, Code : 36 E-Mail : accounts@mirrant.com		Invoice No. 2316 e-Way Bill No. 111961834837 Dated 18-Oct-24				
Delivery Note		Mode/Terms of Payment				
Reference No. & Date.		Other References				
Consignee (Ship to) GV One Plot. No. 15-B, Sy. No. 230 to 243, Turkapally village, Shamirpet mandal, Medchal Malkajgiri District, 500078 GSTIN/UID : 36AADCB2608M1ZO State Name : Telangana, Code : 36		Buyer's Order No. 20240926006 Dated 26-Sep-24 Dispatch Doc No. Delivery Note Date Dispatched through Destination				
Buyer (Bill to) CRESCENTIA LABS PRIVATE LIMITED 5-4-187/3 & 4, M G Road, Secunderabad -500003, 81215 51388, . GSTIN/UID : 36AADCB2608M1ZO State Name : Telangana, Code : 36		Terms of Delivery				
SI No.	Description of Goods	HSN/SAC	Quantity	Rate	per	Amount
1	80MM ELECTROMAGNETIC BTU METER MAKE:MIRRANT; MODEL:BTU 2900 SR NO : M922102024	90261010	1 NO'S	46,200.00	NO'S	46,200.00
	CGST					4,158.00

continued to page number 2

This is a Computer Generated Invoice

INWARD	
Inward No: 7049	Dr: 18/10/24
MRN No:	Dr:
Received by:	Sign:
CRESCENTIA LABS PVT LTD	

MRN :- 20241019006

Tax Invoice(Page 2)

MIRANT AUTOMATION PRIVATE LIMITED PLOT NO:6A/5, IDA PHASE-I, PATANCHERU, HYDERABAD, TS (36) 502319 GSTIN/UIN: 36AAKCM0817G1ZK State Name : Telangana, Code : 36 E-Mail : accounts@mirrant.com	Invoice No. 2316 Delivery Note Reference No. & Date.	e-Way Bill No. 111961834837 Mode/Terms of Payment Other References	Dated 18-Oct-24 Buyer's Order No. 20240926006 Dispatch Doc No. Dispatched through Destination
Consignee (Ship to) GV One Plot. No. 15-B, Sy. No. 230 to 243, Turkapally village, Shamirpet mandal, Medchal Malkajgiri District, 500078 GSTIN/UIN : 36AADCB2608M1ZO State Name : Telangana, Code : 36	Terms of Delivery		
Buyer (Bill to) CRESCENTIA LABS PRIVATE LIMITED 5-4-187/ 3 & 4, M G Road, Secunderabad -500003, 81215 51388, . GSTIN/UIN : 36AADCB2608M1ZO State Name : Telangana, Code : 36			

SI No.	Description of Goods	HSN/SAC	Quantity	Rate	per	Amount
	SGST					4,158.00
Total			1 NO'S			₹ 54,516.00

Amount Chargeable (in words)

INR Fifty Four Thousand Five Hundred Sixteen Only

E. & O.E

HSN/SAC	Taxable Value	Central Tax		State Tax		Total Tax Amount
		Rate	Amount	Rate	Amount	
90261010	46,200.00	9%	4,158.00	9%	4,158.00	8,316.00
Total	46,200.00		4,158.00		4,158.00	8,316.00

Tax Amount (in words) : **INR Eight Thousand Three Hundred Sixteen Only**

Company's Bank Details
 A/c Holder's Name: **Mirrant Automation Private Limited**
 Bank Name : **ICICI BANK**
 A/c No. : **180605000774**
 Branch & IFS Code: **PATANCHERU & ICICI0007318**

Company's PAN : **AAKCM0817G**

Declaration

We declare that this invoice shows the actual price of the goods described and that all particulars are true and correct.

for MIRANT AUTOMATION PRIVATE LIMITED

Authorised Signatory

This is a Computer Generated Invoice



e-Way Bill

e-Way Bill

Doc No. : Tax Invoice - 2316
Date : 18-Oct-24

IRN : b0964b1848159e0ba1ebbf8a7f2b31b8929fcbbbe1d765f01df00eaa8efa74ba
Ack No. : 112422285111059
Ack Date : 18-Oct-24

**1. e-Way Bill Details**

e-Way Bill No. : 111961834837 Mode : Generated Date : 18-Oct-24 11:58 AM
Generated By : 36AAKCM0817G1ZK Approx Distance : 40 KM Valid Upto :
Supply Type : Outward-Supply Transaction Type : Bill To - Ship To

2. Address Details**From**

MIRANT AUTOMATION PRIVATE LIMITED
GSTIN : 36AAKCM0817G1ZK
Telangana

To

CRESCENTIA LABS PRIVATE LIMITED
GSTIN : 36AADCB2608M1ZO
Telangana

Dispatch From

PLOT NO:6A/5, IDA PHASE-I, PATANCHERU,,
HYDERABAD, TS (36) 502319
hyderabad Telangana 502319

Ship To

Plot. No. 15-B, Sy. No. 230 to 243, Turkapally village,,
Shamirpet mandal, Medchal Malkajgiri District,, 500078
Turkapally Telangana 500078

3. Goods Details

HSN Code	Product Name & Desc	Quantity	Taxable Amt	Tax Rate (C+S)
90261010	80MM ELECTROMAGNETIC BTU METER & 80MM ELECTROMAGNETIC BTU METER	1 NOS	46,200.00	9+9

Tot.Taxable Amt : 46,200.00 Other Amt : Total Inv Amt : 54,516.00
CGST Amt : 4,158.00 SGST Amt : 4,158.00

4. Transportation Details

Transporter ID : 36AAKCM0817G1ZK
Name : MIRANT AUTOMATION PRIVATE LIMITED

Doc No. :
Date :

5. Vehicle Details

Vehicle No. : From : hyderabad

CEWB No.:





MIRRANT

Mirrart Automation Pvt.Ltd.
Plot No:6A/5,IDA Phase-I, Patancheru,Sangareddy Dist.,
Hyderabad-502319,Telangana,INDIA

Calibration Certificate

Certificate No : MAPL/MAGFLOW/2024	O.A.No : MAPL/SAL24-25 / M922102024
Calibration Date : 15-10-2024	Due Date : 14-10-2025
Product : MagFlow 2400	Sr No : M922102024
Customer Name : Crescentia Labs Pvt Ltd	

Details of Flow meter

Type of Flowmeter :	Electromagnetic /BTU	Power Supply :	230 V AC
Line Size :	DN 80	Accuracy :	(+/-) 0.5% of measured value
Calibration Factor:	0.21166491	Output Signal:	4 - 20 mA DC
Calibrated Range :	0 - 100 m3/h	Calibrated Range :	0 - 100 m3/h

List of Formulae

$$1) Q_c = \frac{(C-4) \times \text{Calibration range}}{16} \quad 2) \% \text{ of Deviation} = \frac{(Q_c - Q_m) \times 100}{Q_m}$$

In Q_c Q_m

Nomenclature

Q_m = Master Flow Rate Q_c = Flow Rate Calculated from
 C = Current output of UUC Current c

Calibration Method : Comparison Method

Traceability : All the Instrument used or traceable to National standards through reference Standards and their calibrations are valid.

Calibration Results

Sr.No	Master Flow Rate (Q_m)	Calculated Flow Rate (Q_c)	Deviation In (Q_c)
	m3/h	m3/h	%
1	25.703	25.724	0.082
2	50.666	50.681	0.030
3	75.111	75.252	0.187

Calibrated By :- HARI

Verified By :- TARIK

Signature :

Hari

Signature :



It is hereby certified that the equipment mentioned above has been tested and found to meet its specifications. The Performance of the above equipment is guaranteed for period of 12 months from the date of installation or 18 months from the date of dispatch, whichever is earlier, against any manufacturing defects only.